



DOLT-3

Dogfish Liver Certified Reference Material for Trace Metals

The following table shows those elements for which certified values have been established for this dogfish (*Squalus acanthias*) liver reference material. Certified values are based on unweighted mean results from data submitted by laboratories participating in an annual intercomparison. The expanded uncertainty (U_{CRM}) in the certified value is equal to $U = k u_c$ where u_c is the combined standard uncertainty calculated according to the ISO Guide [1] and k is the coverage factor. The value of u_c is determined from the combined uncertainties of the various methods (u_{char}) as well as uncertainties associated with homogeneity (u_{hom}).

It is intended that U_{CRM} encompasses every aspect that reasonably contributes to the uncertainty of the measurand [2,3]. A coverage factor of 2 was applied for all elements. The table below lists certified values for DOLT-3.

TRACE ELEMENTS (milligram/kilogram)

Arsenic (d,g,h)	10.2	±	0.5
Cadmium (d,f,g,i,p)	19.4	±	0.6
Copper (d,i,p)	31.2	±	1.0
Iron (d,f,i)	1484	±	57
Lead (d,g,p)	0.319	±	0.045
Mercury (d,c,p)	3.37	±	0.14
Nickel (g,i,p)	2.72	±	0.35
Selenium (g,h,j)	7.06	±	0.48
Silver (f,g,i,p)	1.20	±	0.07
Zinc (d,i,p)	86.6	±	2.4

Coding

The coding refers only to the instrumental method of analyte determination.

- | | |
|---|--|
| c - Cold vapour atomic absorption spectrometry. | h - Hydride generation atomic absorption spectrometry. |
| d - Inductively coupled plasma mass spectrometry | i - Inductively coupled plasma atomic emission spectrometry. |
| f - Flame atomic absorption spectrometry. | j - Hydride generation atomic fluorescence spectrometry. |
| g - Electrothermal vaporization atomic absorption spectrometry (ETAAS). | p - Isotope dilution inductively coupled plasma mass spectrometry. |

Intended Use

This reference material is primarily intended for use in the calibration of procedures and the development of methods used for the analysis of marine fauna and materials with a similar matrix.

Storage and Sampling

This material should be kept tightly closed in the original bottle and should be stored in a cool location, away from any significant radiation sources such as ultraviolet lamps and sunlight. The bottle should be well mixed by rotation and shaking prior to use, and tightly closed immediately thereafter.

Instructions for Drying

DOLT-3 can be dried to constant weight by:

- (1) drying at reduced pressure (e.g., 50 mm Hg) at room temperature in a vacuum desiccator over magnesium perchlorate for 24 hours;
- (2) vacuum drying (about 0.5 mm Hg) at room temperature for 24 hours.

Stability


The predecessor CRM, DOLT-2, has been periodically analyzed for more than nine years and found to be both physically and chemically stable over this time interval. We expect similar behaviour from DOLT-3. The stability of this CRM will continue to be monitored and customers will be notified if any significant irregularity occurs prior to the expiry date.

Expiry

Based on sample stability noted above, the certified values for DOLT-3 are considered valid until September, 2011, provided the CRM is handled and stored in accordance with instructions herein.

Preparation of DOLT-3

This reference material was processed at the Guelph Food Technology Center, Guelph Ontario. The preparation sequence is illustrated below.



Information values

Due to the scatter of results, a certified value for Cr was not calculated. A lack of independent values precluded the certification of Sn and methylmercury. Information values for these analytes are given below.

Al	25	mg/kg
Cr	3.5	mg/kg
Sn	0.4	mg/kg
CH ₃ Hg (as Hg)	1.7	mg/kg

Certified value

DOLT-3 was provided as an unknown sample to a group of laboratories participating in an annual intercomparison for trace metals in marine samples sponsored by NRC [4]. The predecessor CRM DOLT-2 was provided as a quality control sample. Data generated at NRC was also included in the pool of intercomparison results.

Laboratories were requested to provide triplicate dry weight values using an analytical method of choice based on total digestion of the sample. Various dissolution procedures were utilized including closed vessel with microwave heating and open vessel digestion on a hot plate. Laboratories used HNO₃ or a combination of H₂O₂ and HNO₃, although several participants included HF or HCl in the acid mixture.

Data was returned to NRC for evaluation. The results from a select sub-group of participants were used for the certification of DOLT-3. These laboratories were selected based on their performance history in previous intercomparisons. As well, if a laboratory did not submit satisfactory results for DOLT-2 their data for DOLT-3 were rejected.

The certified values were calculated from the unweighted means of the results of the participating laboratories. [5,6] Data were first examined for outliers using the Grubb's Test; no outliers of means were detected.

New guidelines for CRM producers suggest all sources relevant to the user of the material should contribute to the uncertainty of the certified value [2-4]. Included in the overall uncertainty estimate are uncertainties in the batch characterisation (μ_{char}), uncertainties related to possible between-bottle variation (μ_{hom}) as well as instability derived from effects relating to long-term storage and transport (μ_{stab}). Expressed as standard uncertainties these components can be combined as:

$$\mu_{c(CRM)}^2 = \mu_{char}^2 + \mu_{hom}^2 + \mu_{stab}^2 \quad (1)$$

Results for the various statistics used to calculate the certified values are shown in Table 2.

Characterisation

The characterisation uncertainties (μ_{char}) were calculated in accordance with equation 2, where s is the standard deviation of the means and p is the number of mean results included in the calculation [6].

$$\mu_{char} = \frac{s}{\sqrt{p}} \quad (2)$$

Homogeneity

The homogeneity components of the uncertainty in the certified values were derived according to the recommendation of an international study group [7]. The material was tested for homogeneity at NRC using ETAAS. Results from triplicate sub-samples (0.150 g) from ten bottles were evaluated using ANOVA.

For Cd, Ni, Al, Pb and Zn, the inhomogeneity contribution to uncertainty, μ_{hom} , was set to the experimentally determined between-unit standard deviation ($s_{between}$), as the best estimate of the uncertainty due to between-unit heterogeneity. However, for As, Cu, Fe, Hg, Se and Ag, the situation depicted in equation 3 occurred:

$$s_{between}^2 < \frac{s_{meas}^2}{n} \quad (3)$$

where s_{meas} is the repeatability standard deviation for the method used in the homogeneity assessment and n is the number of replicates per unit.

In these cases, μ_{hom} was calculated according to:

$$\mu_{hom} = \sqrt{\frac{s_{meas}}{n}} \quad (4)$$

It is recognized that this is not an ideal situation, as it represents a worst case scenario by suggesting the homogeneity could be as poor as the precision of the measurement technique selected for homogeneity assessment.

Stability

As previously discussed, based on NRC's experience with similar materials, uncertainty components for long and short term stability were considered negligible and are thus not included in the uncertainty budget.

Table 2. Statistical Data for DOLT-3

	components for μ_{char} , (mg/kg)			data sets	μ_{char} , (mg/kg)	μ_{hom} , (mg/kg)
	S_L	S_w	s of means			
As	0.62	0.26	0.64	12	0.18	0.14
Cd	0.8	0.34	0.82	16	0.21	0.23
Cu	1.59	0.78	1.63	16	0.41	0.24
Fe	59	26	61	13	17	23
Pb	0.036	0.015	0.037	12	0.011	0.02
Hg	0.149	0.09	0.158	14	0.042	0.058
Ni	0.27	0.15	0.29	13	0.08	0.08
Se	0.59	0.44	0.64	9	0.22	0.11
Ag	0.1	0.07	0.107	13	0.029	0.016
Zn	3.9	1.3	3.9	15	1	0.6

s_L - the between laboratory standard deviation from the laboratory intercomparison

s_w - the within laboratory standard deviation from the laboratory intercomparison

Criteria for acceptance

The property values reported in this certificate are the best estimates of the true values that can be obtained, based on the certification exercise. It is the responsibility of the analyst to assess the appropriateness of this CRM for purpose and interpret their own analytical results [8].

The user may assess laboratory bias from the difference between the calculated mean value of replicate measurements (\bar{x}) and the certified value (μ): $\bar{x} - \mu$. According to ISO Guide 33: Uses of Certified Reference Materials, the criteria for acceptance is:

$$-a_2 - 2\sigma_D \leq \bar{x} - \mu \leq a_1 + 2\sigma_D \quad (5)$$

where a_1 and a_2 are adjustment values chosen by the laboratory according to economic or technical limitations or requirements and σ_D is the standard

deviation associated with the measurement process. The value of σ_D can be calculated from the user's laboratory quality control data and may be estimated from two components, as shown in equation 6:

$$\sigma_D = \sqrt{s_b^2 + \frac{s_w^2}{n}} \quad (6)$$

where s_b is the between-laboratory standard deviation component associated with the measurement process (ideally this should include the long-term standard deviation of the user's method or, alternatively, s_L from Table 2), s_w the within-laboratory standard deviation (or repeatability standard deviation) and n the number of replicate analyses made of the reference material. Alternative methods for calculating s_b are discussed in reference 8.

Acknowledgements

The following staff members of the Institute for National Measurement Standards, National Research Council of Canada, participated in the certification: V.J. Boyko, C. Scriver, P. Maxwell, L. Yang and S. Willie.

The cooperation of I. Britt and A. Mannen of the Guelph Food Technology Centre, Guelph, ON, Canada in the preparation of this material is gratefully acknowledged.

The following laboratories participated in the certification of DOLT-3:

ALS Environmental
1988 Triumph St.
Vancouver, B.C. V5L 1K5

Australian Institute of Marine Sciences
Townsville Mail Center
Queensland, 4810 , Australia
Frank Tirendi, Cassie Payn and Stephen Boyle

Battelle Pacific Northwest
1529 W. Sequim Bay Road
Sequim, WA 98382
Eric Crecelius, Laurie Niewolny, Mary Ann Deuth,
and Brenda Lasorsa

CanTest Ltd.
4606 Canada Way
Burnaby, B.C. V5G 1K5
Richard Jorntitz, Lynn Yehle, Alex Wong and
Greg Sparrow

Massachusetts Water Resources Authority
Central Laboratory
190 Taft Ave.
Winthrop, MA 02152
Patricia Sullivan , Maria Palumbo, Jun Wang,
Ellen C. West and Larissa Zilitinkevitch

NOAA
National Marine Fisheries Service
Charleston Laboratory
217 Ft. Johnson Road
Charleston, SC 29412
Dan Bearden, Aaron Dias

Skidaway Institute of Oceanography
10 Ocean Science Circle
Savannah, GA 31411
Herb Windom, Debbie Wells and Lori Cowden

Texas A. & M.
College of Veterinary Medicine
Trace Element Research Laboratory
College Station, TX 77843-4458
Robert Taylor, Gerald Bratton and Bryan Brattin

Texas Parks and Wildlife
Environmental Contaminants Laboratory
505 Staples Road
San Marcos, TX 78666
Gary Steinmetz

U.S. Customs Laboratory
214 Bourne Boulevard
Savannah, GA 31408
Ralph Smith

USGS
National Water Quality Laboratory
P.O. Box 25046
Denver, Colorado 80225-0046

USGS-WRD
Peachtree Business Center
3039 Amwiler Road
Atlanta, GA 30360-2824
Kent Elrick, Arthur J. Horowitz and James J.
Smetanka

WA State Dept. of Ecology
Manchester Environmental Laboratory
7411 Beach Drive East
Port Orchard, Washington 98366-8204
Randy Knox, Meredith Jones, Sally Cull and
Karin Feddersen

References

- [1] Guide to the Expression of Uncertainty in Measurement, ISBN 92-67-10188-9, 1st ed. ISO, Geneva, Switzerland (1993).
- [2] J. Pauwels, A. van der Veen, A. Lamberty, H. Schimmel, Evaluation of uncertainty of reference materials. *Accred Qual Assur* (2000) 5:95-99.
- [3] J. Pauwels, A. Lamberty, H. Schimmel, The determination of the uncertainty of reference materials certified by laboratory intercomparison. *Accred Qual Assur* (1998) 3:180-184.
- [4] S. Willie, Fifteenth Intercomparison for Trace Elements in Marine Sediments and Biological Tissues, NRC No. 42768, May 2002.
- [5] A. M.H.van der Veen and J. Pauwels, Uncertainty calculations in the certification of reference materials. 1.Principles of analysis of variance. *Accred Qual Assur* (2000) 5:464-469.
- [6] A. M.H. van der Veen, T. P.J. Linsinger, H. Schimmel, A. Lamberty and J. Pauwels, Uncertainty calculations in the certification of reference materials 4. Characterisation and certification. *Accred Qual Assur* (2001) 6:290-294.
- [7] S.L.R. Ellison, S.Burke, R.F.Walker, K. Heydorn, M.Månsson, J.Pauwels, W.Wegscheider, B.te Nijenhuis, Uncertainty for reference materials certified by interlaboratory study: Recommendations of an international study group. *Accred Qual Assur* (2001) 6:274-277.
- [8] R. Walker and I. Lumley, Pitfalls in terminology and use of reference materials, *Trends in Anal. Chem.*, (1999) 18: 594-616.

Updates

It is anticipated that as more data become available, the established values may be updated and reliable values assigned to more elements.

Our web site at <http://www.cm.inms.nrc.ca> will contain any new information.

Certificate issued September 2002.

Visit our web site at
<http://www.cm.inms.nrc.ca>

Comments, information and inquiries should be addressed to:

Dr. R.E. Sturgeon
National Research Council of Canada
Institute for National Measurement Standards
M-12, 1500 Montreal Road
Ottawa, Ontario, Canada K1A 0R6

Telephone (613) 993-2359
Facsimile (613) 993-2451
E-mail crm.inms@nrc.ca

Également disponible en français sur demande.