



## Taiga Environmental Laboratory

4601 52<sup>nd</sup> Avenue – Yellowknife, NT X1A 2R3

Phone: (867) 669-2788 Fax: (867) 669-2718 Email: taiga@inac-ainc.gc.ca

### Water Sampling Instructions

### Collecting the Sample



#### Step One:

Prior to sampling, ensure you have obtained all the sampling equipment you require, such as the proper bottles, filtration devices, *etc.* Refer to the Taiga's Water Sampling Instructions – Ordering Bottles. If there are any questions or concerns, do not hesitate to contact the laboratory. Please have your water license (if applicable) available before contacting the laboratory to ensure proper bottles are ordered. **Note: you may need more than one bottle per sampling site.**



#### Step Two:

Check your local departure flight schedule to Yellowknife for the day you plan to take your samples. Samples should be shipped to the Laboratory **as soon as possible** after collection. Time your sampling so that the samples can be shipped out by plane as soon as possible.



#### Step Three:

Follow the sampling instructions on the back of this sheet for each bottle type. Package bottles in a cooler and send to the laboratory. If you require microbiological tests, such as Total Coliforms, E. coli., Fecal Coliforms, BOD, *etc.*, please contact the laboratory with the collection date and time, the Airline name, the waybill number and the expected time of arrival.



#### Safety Issues:

Wear appropriate gloves when collecting any sample to avoid contamination and possible exposure to unhealthy substances. The sample preservatives provided by the Laboratory are corrosive and will cause a burning sensation on the skin. If you should spill any on your skin or clothes, rinse the area **immediately** with lots of cool water. Call a doctor should the burning sensation continue.



# Taiga Environmental Laboratory

4601 52<sup>nd</sup> Avenue – Yellowknife, NT X1A 2R3

Phone: (867) 669-2788 Fax: (867) 669-2718 Email: taiga@inac-ainc.gc.ca

## Water Sampling Instructions

## Collecting the Sample

Parameter Group		Marking	Preservative	Instructions
●	Routine	GREEN	Keep cool at 4°C	1. Rinse bottle three (3) times with sample 2. Fill to top and cap bottle.
●	Nutrients	BLACK	Keep cool at 4°C	
●	Biochemical Oxygen Demand (BOD)	PURPLE	Keep cool at 4°C	1. Rinse bottle three (3) times with sample 2. Fill to top and cap bottle. 3. <b>Sample must be sent to laboratory within 24 hours</b>
	Microbiological	STERILE	Sodium Thiosulphate and Keep cool at 4°C	1. <b>DO NOT RINSE BOTTLE</b> 2. Fill to top and cap bottle. 3. <b>Sample must be sent to laboratory within 24 hours</b>
●	Total Metals	RED	5 mL of 1:3 nitric acid in RED-dot vials	1. Rinse bottle three (3) times with sample 2. Fill to near the top. 3. Add contents of preservative vial 4. Cap bottle and mix.
●	Dissolved Metals	RED	5 mL of 1:3 nitric acid in RED-dot vials	1. <b>Filter Sample with 45 µm filter</b> 2. Rinse bottle three (3) times with <b>filtrate</b> 3. Fill to near the top. 4. Add contents of preservative vial 5. Cap bottle and mix.
●	Hexane Extractable Material (HEM) (also known as Oil and Grease)	YELLOW	4 mL 1:1 sulphuric acid in YELLOW-dot vial	1. <b>DO NOT RINSE BOTTLE</b> 2. Fill to shoulder of bottle. 3. Add contents of preservative vial 4. Cap bottle and mix
	BTEX, THM & Purgeable Hydrocarbons	40 mL CLEAR GLASS W/ WHITE LID	Keep cool at 4°C	1. <b>DO NOT RINSE BOTTLE</b> 2. Fill bottle completely leaving <b>NO</b> air bubbles
	Extractable Hydrocarbons	1 L AMBER GLASS W/ BLACK LID	Keep cool at 4°C	1. <b>DO NOT RINSE BOTTLE</b> 2. Fill to top and cap bottle.