

National Research Council Canada Conseil national de recherches Canada

MOOS-1

Seawater Certified Reference Material for Nutrients

This reference material is primarily intended for use in the calibration of procedures and the development of methods used for the analysis of nutrients in seawater.

The following table shows those elements for which certified values have been established for MOOS-1.

Nutrient (micromole/liter)

Orthophosphate	1.56	±	0.07
Silicate	26.0	±	1.0
Nitrite	3.06	±	0.15
Nitrite and Nitrate	23.7	±	0.9

Certified values are based on unweighted mean results from data submitted by laboratories participating in an annual intercomparison for the determination of nutrients, as well as results from NRC obtained by two independent analytical methods. The expanded uncertainty (U_{CRM}) in the certified value is equal to U = ku_c where u_c is the combined standard uncertainty calculated according to the ISO Guide [1] and k is the coverage factor. The value of u_c is determined from the combined uncertainties of the various methods (u_{char}) as well as uncertainties associated with homogeneity (u_{hom}) and stability (u_{stab}) .

It is intended that U_{CRM} encompass every aspect that reasonably contributes to the uncertainty of the measurand [2,3]. A coverage factor of 2 was applied for all analytes.



Storage and Sampling

This material should be stored in the original bottle in a refrigerator (4°C), although samples stored at $+20^{\circ}$ C or $+40^{\circ}$ C for two month periods have not undergone detectable changes.

Once opened, contamination may cause changes in the analytes. Care should be taken if it is intended to remove only a portion of the sample and store the remainder for use at a later date. The bottle should be opened for the minimal period in a clean area and otherwise remain tightly closed.

Collection of Water for MOOS-1

MOOS-1 was collected at Lat. 47.062833 °N, Long. 59.982333 °W, off the northern tip of Cape Breton Island, NS, Canada. The water was collected from a depth of about 200 meters using a rosette containing 22 Niskins, each of about 10L volume. The contents of each Niskin were transferred, using a peristaltic pump, through a 0.05 µm cartridge filter into 50L carboys. The water was homogenized in a 400L tank, bottled in 50 mL aliquots in 60 mL bottles, sealed, and irradiated. The water was collected June 24, bottled July 11 & 12, and irradiated July 16, 1996. It was later found that irradiation caused bottle to bottle inhomogeniety and a portion of the original sample was reblended and bottled on April 3, 2001, without subsequent irradiation.

Certified value

Guidelines for CRM producers suggest all sources relevant to the user of the material contribute to the uncertainty of the certified value [2,3]. Included in the overall uncertainty estimate are uncertainties associated with the batch characterisation (u_{char}) , uncertainties related to possible between-bottle variation (u_{hom}) as well as instability derived from effects relating to long-term storage and short term transport (u_{stah}) . Expressed as standard

uncertainties, these components can be combined as:

$$u_{\rm c(CRM)}^2 = u_{\rm char}^2 + u_{\rm hom}^2 + u_{\rm stab}^2$$
 (1)

Results for the various statistics used to calculate the certified values are summarized in Table 2.

Characterisation

MOOS-1 was analysed by a number of expert laboratories participating in an annual intercomparison for nutrients sponsored by the National Oceanic and Atmospheric Administration (NOAA) [4]. Data generated at NRC using two independent methods of analysis were also included in the pool of intercomparison results.

Laboratories were requested to provide duplicate results using an analytical method of their choice.

Data were returned to NRC for evaluation. The results from a select sub-group of participants were used for the certification of MOOS-1. These laboratories were selected based on their satisfactory performance history in a previous intercomparison.

The certified values were calculated from the unweighted means of the results of the participating laboratories [3,5-7]. Data were first examined for outliers using the Grubb's Test.

The characterisation uncertainties (u_{char}) were calculated in accordance with Equation 2, where s is the standard deviation of the means and *p* is the number of mean results included in the calculation [6].

$$u_{\rm char} = \frac{s}{\sqrt{p}}$$
 (2)

Table 2. Statistical Data for MOOS-1

Uncertainty components, micromole/liter	Orthophosphate	Silicate	Nitrite	Nitrite + Nitrate
from				
intercomparison				
SL	0.096	1.6	0.116	0.57
S _w	0.011	0.20	0.025	0.21
s of means	0.096	1.6	0.118	0.59
data sets, p	14	14	14	14
U _{char}	0.026	0.43	0.032	0.16
U _{hom}	0.011	0.17	0.029	0.31
U _{stab}	0.018	0.18	0.054	0.28

s₁ - the between laboratory standard deviation from the laboratory intercomparison

 s_{w}^{-} - the within laboratory standard deviation from the laboratory intercomparison

Analytical Methods

For the determination of nutrients in MOOS-1, laboratories predominantly used methods based on the traditional colorimetric procedures of Strickland and Parsons [8] and Morris and Riley [9]. Recently, automated colorimetric methods have been prepared by the USEPA based on these principles [10-12].

Independent methods based on ion chromatrography for nitrite and nitrate, and ion exclusion chromatography inductively coupled plasma mass spectrometry for silicate and phosphate were developed at NRC to corroborate the colorimetric results.

Homogeneity

The homogeneity components of the uncertainty in the certified values were derived according to the recommendation of an international study group [7]. The material was tested for homogeneity at NRC using standard colorimetric procedures. Results from triplicate sub-samples from ten bottles were evaluated using ANOVA.

The inhomogeneity contribution to uncertainty, u_{hom} , was set equal to the experimentally determined between-unit standard deviation (s_{between}), as the best estimate of the uncertainty due to homogeneity.

Stability

Uncertainty components for long and short term stability were evaluated. To determine possible uncertainty associated with transport or short term instability, samples of MOOS-1 were stored at +40°C, +20°C and -25°C for one month. These samples were analysed at the same time as several bottles which had been stored under recommended conditions of +4°C. No between bottle differences were observed, therefore the uncertainty component for short term stability was set to zero.

The uncertainty components related to the long term stability of this CRM were calculated according to the recommendations of an international study group [7]. The slope and uncertainty in the regression fit of the temporal stability data were used to calculate the values in Table 2, based on a projected 60 month lifetime.

MOOS-1 has been repetitively monitored since 1996 and found to be stable with respect to nutrient concentration over this period. The stability of this CRM will continue to be monitored and customers will be notified if any significant irregularity occurs prior to the expiry date.

Expiry

Based on sample stability, the certified values for MOOS-1 are considered valid until December 2008, provided the CRM is handled and stored in accordance with instructions herein.

Criteria for acceptance

The property values reported in this certificate are the best estimates of the true values that can be obtained, based on the certification exercise. It is the responsibility of the analyst to assess the appropriateness for purpose of this CRM and interpret their own analytical results [13].

The user may assess laboratory bias from the difference between the calculated mean value of replicate measurements (x) and the certified value (μ): x- μ . According to ISO Guide 33: Uses of Certified Reference Materials, the criteria for acceptance is:

$$-a_2 - 2\sigma_D \le \bar{x} - \mu \le a_1 + 2\sigma_D$$
 (5)

where a_1 and a_2 are adjustment values chosen by the laboratory according to economic or technical limitations or requirements and σ_D is the standard deviation associated with the measurement process. The value of σ_D can be calculated from the user's laboratory quality control data and may be estimated from two components, as shown in equation 6:

$$\sigma_D = \sqrt{s_b^2 + \frac{s_w^2}{n}}$$
 (6)

where s_{b} is the between-laboratory standard deviation component associated with the measurement process (ideally this should include the long-term standard deviation of the user's method or, alternatively, s_{L} from Table 2), s_{w} is the within-laboratory standard deviation (or repeatability standard deviation) and n is the number of replicate analyses made of the reference material. Alternative methods for calculating s_{b} are discussed in reference 13.

Acknowledgements

The following staff members of the Institute for National Measurement Standards, National Research Council of Canada, participated in the certification: V.P. Clancy, J.W. McLaren, L. Yang and S. Willie.

The cooperation and advice of Dr. Peter Strain (Bedford Institute of Oceanography, Dartmouth, Nova Scotia) in the preparation of this material is gratefully acknowledged.

NOAA's Center for Coastal Monitoring and Assessment, Silver Spring, MD for sponsorship of the intercomparison.

The following laboratories participated in the certification of MOOS-1:

Fisheries and Oceans Canada Bedford Institute of Oceanography Marine Nutrient Lab Dartmouth, N.S. B2Y 4A2 Peter Strain, Pierre Clement and Carol Anstey

Fisheries and Oceans Canada Institute of Ocean Sciences 9860 West Saanich Rd. Sidney, B.C. V8L 4B2 Janet Barwell Clarke

Fisheries and Oceans Canada Institut Maurice Lamontagne 850, route de la Mer, C.P. 1000 MontJoli, Québec, Canada Alain Gagné

King County Environmental Laboratory Conventionals Unit 322 West Ewing St Seattle, WA 98119 Despina Strong

Louisiana Universities Marine Consortium 8124 Highway 56 Chauvin, LA 70344 Grady Hanrahan, Rodney Powell, Sheryl Bell and Amy Wilson-Finelli. Miami Dade DERM Laboratory 211 West Flagler Street Miami, FL 33130 Jana Bares, Dora Magrisso and Julio Lopez

NOAA

4301 Rickenbacker Causeway Miami, FL 33149 Jia-Zhong Zhang

U.S. EPA Atlantic Ecology Division 27 Tarzwell Dr. Narragansett, RI 02882 Marnita Chintala

University of California Marine Science Institute Santa Barbara, CA 93106 Robert Petty, William Clinton and Anya Engen

University of Connecticut Environmental Research Institute Storrs, CT 062695210 Christopher Perkins, Stephanie Baver and Elizabeth Fumal

University of Louisiana at Lafayette Center for Ecology and Environmental Technology Lafayette, LA 70504 Robert Twilley

University of Plymouth School of Environmental Sciences, Plymouth Environmental Research Centre, Plymouth, PL48AA UK Paul Worsfold, Paulo Gardolinski and Utra Mankasingh

University of Washington School of Oceanography Seattle, WA 981957940 Katherine Krogslund

Scripps Institution of Oceanography 8855 Biological Grade La Jolla, CA 92037 Doug Masten

References

[1] Guide to the Expression of Uncertainty in Measurement, ISBN 92-67-10188-9, 1st ed. ISO, Geneva, Switzerland (1993).

[2] J. Pauwels, A. van der Veen, A. Lamberty, H. Schimmel, Evaluation of uncertainty of reference materials.

Accred Qual Assur (2000) 5:95-99.

[3] J. Pauwels, A. Lamberty, H. Schimmel, The determination of the uncertainty of reference materials certified by laboratory intercomparison. Accred Qual Assur (1998) 3:180-184.

[4] S.N. Willie and V.P. Clancy, Second Intercomparison for Nutrients in Seawater, NOAA Technical Memorandum 158, October 2002.

[5] A.M.H.van der Veen and J. Pauwels, Uncertainty calculations in the certification of reference materials. 1.Principles of analysis of variance.

Accred Qual Assur (2000) 5:464-469.

[6] A.M.H. van der Veen, T.P.J. Linsinger,
H.Schimmel, A. Lamberty and J. Pauwels,
Uncertainty calculations in the certification of reference materials 4. Characterisation and certification.

Accred Qual Assur (2001) 6:290-294.

[7] S.L.R. Ellison, S.Burke, R.F.Walker, K. Heydorn, M.Månsson, J.Pauwels, W.Wegscheider, B.te Nijenhuis, Uncertainty for reference materials certified by interlaboratory study: Recommendations of an international study group. Accred Qual Assur (2001) 6:274–277.

[8] J.D. Strickland and T. R. Parsons. 1972. A practical handbook of seawater analysis. Bull. Fish. Res. Board Can. 167: 310p.

[9] A.W. Morris and J.P. Riley, Determination of nitrate in sea water. Anal. Chim. Acta. (1963) 29:272-279.

[10] J. Zhang and G. Berberian, EPA Method
366.0, Determination of dissolved silicate in estuarine and coastal waters by gas segmented continuous flow colorimetric analysis. Sept., 1997.
U.S. Environmental Protection Agency, Cincinnati, Ohio.

[11] C.F. Zimmermann and C.W. Keefe, EPA Method 365.5, Determination of orthophosphate in estuarine and coastal waters by automated colorimetric analysis. Sept., 1997. U.S. Environmental Protection Agency, Cincinnati, Ohio.

[12] J. Zhang, P.B.Ortner and C.J. Fischer, EPA Method 353.4, Determination of Nitrate and Nitrite in Estuarine and Coastal Waters by Gas Segmented Continuous Flow Colorimetric Analysis. Sept., 1997. U.S. Environmental Protection Agency, Cincinnati, Ohio.

[13] R. Walker and I. Lumley, Pitfalls in terminology and use of reference materials, Trends in Anal. Chem., (1999) 18: 594-616.

Updates

Our web site will contain any new information. http://inms-ienm.nrc-cnrc.gc.ca/calserv/ chemical_metrology_e.html

> Certificate issued March 2003. Expiry date revised November 2007.

Comments, information and inquiries should be addressed to:

Dr. R.E. Sturgeon National Research Council of Canada Institute for National Measurement Standards M-12, 1500 Montreal Road Ottawa, Ontario, Canada K1A 0R6

Telephone Facsimile E-mail (613) 993-2359 (613) 993-2451 crm.inms@nrc-cnrc.gc.ca

Également disponible en français sur demande.

Canadä